

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 03/03/04

Date Received: 02/12/04

Project: Metro Self Monitor, PO# M92793, F&BI 402094

Date Extracted: 03/02/04

Date Analyzed: 03/02/04

**RESULTS FROM THE ANALYSIS OF THE WATER SAMPLE
FOR TOTAL METALS BY
INDUCTIVELY COUPLED PLASMA (ICP)
(METHOD 6010)**

Results Reported as mg/L (ppm)

<u>Sample ID</u> Laboratory ID	<u>Chromium</u>	<u>Copper</u>	<u>Nickel</u>	<u>Zinc</u>
M92793 402094-01	0.16	0.19	0.17	<0.05
Method Blank	<0.05	<0.05	<0.10	<0.05

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 03/03/04

Date Received: 02/12/04

Project: Metro Self Monitor, PO# M92793, F&BI 402094

QUALITY ASSURANCE RESULTS FROM TOTAL METALS BY INDUCTIVELY COUPLED PLASMA (ICP) (METHOD 6010)

Laboratory Code: 402251-01 (Duplicate)

Analyte	Reporting Units	Sample Result	Duplicate Result	Relative Percent Difference	Acceptance Criteria
Chromium	mg/L (ppm)	<0.05	<0.05	nm	0-20
Copper	mg/L (ppm)	0.14	0.14	0	0-20
Nickel	mg/L (ppm)	<0.1	<0.1	nm	0-20
Zinc	mg/L (ppm)	0.25	0.25	0	0-20

Laboratory Code: 402251-01 (Matrix Spike)

Analyte	Reporting Units	Spike Level	Sample Result	% Recovery MS	Acceptance Criteria
Chromium	mg/L (ppm)	2	<0.05	92	80-120
Copper	mg/L (ppm)	2	0.14	95	80-120
Nickel	mg/L (ppm)	4	<0.1	92	80-120
Zinc	mg/L (ppm)	2	0.25	89	80-120

Laboratory Code: Laboratory Control Sample

Analyte	Reporting Units	Spike Level	% Recovery LCS	% Recovery LCSD	Acceptance Criteria	RPD (Limit 20)
Chromium	mg/L (ppm)	2	103	111	80-120	7
Copper	mg/L (ppm)	2	99	110	80-120	11
Nickel	mg/L (ppm)	4	101	110	80-120	9
Zinc	mg/L (ppm)	2	102	110	80-120	8

nm - The analyte was not detected in one or more of the duplicate analyses. Therefore, calculation of the RPD is not applicable.

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

March 3, 2004

DUPLICATE COPY

INVOICE # 04ACU0303-2

Accounts Payable
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134

RE: Project Metro Self Monitor, PO# M92793, F&BI 402094 - Results of testing
requested by Gerry Thompson for material submitted on February 12, 2004.

1 sample analyzed for Total Chromium, Copper, Nickel and Zinc by Method 6010B @ \$80 per sample	\$ 80.00
--	----------

Amount Due	\$ 80.00
------------------	----------

FEDERAL TAX ID #(b) (6)

402094

SAMPLE CHAIN OF CUSTODY

CM 02/12/04

BI,

Send Report To

G. Thompson

Company

ALASKAN COPPER WORKS

Address

628 S. Harvard St

City, State, ZIP

Seattle WA 98134

Phone #

206-382-8378

Fax #

206-382-4309

SAMPLERS (signature)

PROJECT NAME/NO.

METRO SELF monitor

PO #

M 92793

REMARKS

Page # _____ of _____

TURNAROUND TIME

☐ Standard (2 Weeks)☐ RUSH

Rush charges authorized by: _____

SAMPLE DISPOSAL

☐ Dispose after 30 days☐ Return samples☐ Will call with instructions

Sample ID	Lab ID	Date	Time	Sample Type	# of containers	ANALYSES REQUESTED										Notes		
						TPH-Diesel	TPH-Gasoline	BTEX by 8021B	VOCs by 8260	SVOCs by 8270	HFS							
m92283	01	2/12/04	10:20	A20	1								a, b, c, d, e, f, g, h, i, j, k, l, m, n, o, p, q, r, s, t, u, v, w, x, y, z					

Friedman & Bruya, Inc.
3012 16th Avenue West

Seattle, WA 98119-2029

Ph. (206) 285-8282

Fax (206) 283-5044

SIGNATURE

Relinquished by:

Received by:

Relinquished by:

Received by:

PRINT NAME

G. Thompson

Fred Wolf

COMPANY

ACW

F&B, Inc

DATE

2/12/04

2/12/04

TIME

10:10 AM

10:10 AM

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

March 3, 2004

Gerry Thompson, Project Manager
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134


Dear Mr. Thompson:

Included are the results from the testing of material submitted on February 12, 2004 from the Metro Self Monitor, PO# M92793, F&BI 402094 project. There are 2 pages included in this report. Any samples that may remain are currently scheduled for disposal in 30 days. If you would like us to return your samples or arrange for long term storage at our offices, please contact us as soon as possible.

We appreciate this opportunity to be of service to you and hope you will call if you should have any questions.

Sincerely,

FRIEDMAN & BRUYA, INC.



Charlene Morrow
Chemist

Enclosures
ACU0303R.DOC